

Experimental conditions

Typically, 300 μl of sample component at 10-100 μM is required, and about 75 μl of the ligand at 10x higher in molar concentration, both interacting molecules must be in identical buffer/solvent. Lower or higher concentrations might be necessary depending on the strength of binding. Typical experiment takes 45 minutes. Temperature range 2-80°C.